

SafeDon™

Hygiene System

White Nitrile Examination Gloves

Latex Free • Powder Free • Non Sterile



Powder Free Synthetic Examination Gloves

SAFEDON HYGIENE SYSTEM NITRILE WHITE
SAFEDON HYGIENE SYSTEM NITRILE LONG CUFF
SAFEDON HYGIENE SYSTEM NITRILE BLUE
SAFEDON HYGIENE SYSTEM VINYL



Key Benefits:

- Hygienic system for dispensing and donning gloves.
- Individual gloves dispensed Cuff 1st™, to help reduce the risk of cross contamination.
- Fast, easy dispensing saves time.
- Keeps gloves away from contaminated surfaces.
- Latex Free and Powder Free.

The NEW
SafeDon™
Hygiene System

Better Hygiene. Better Safety. Better Economy.

SafeDon™

Hygiene System

White Nitrile Examination Gloves
Latex Free • Powder Free • Non Sterile



A Vital Part of Your Infection Control System

Medical gloves provide protection from unwanted substances for both healthcare professionals and patients. SafeDon Hygiene System Nitrile Gloves meet and exceed international products standards to provide uncompromising protection to its users.

SafeDon Hygiene System

Cross-contamination is a growing concern in any working environment where glove use and adherence to hand hygiene policies are required. Despite the introduction of hand washing initiatives the level of compliance cannot be guaranteed to reach 100 %. International studies continue to show that gloves and traditional glove boxes themselves can be a vehicle for contamination. The new innovative SafeDon Hygiene System dispense only one glove at the time cuff 1st, eliminating touching of the glove most critical surfaces - thumb, finger and palm area - and reduce the risk of cross infections (1-6)

SafeDon Hygiene System is a hygienic initiative which further helps to reduce the risk of cross-contamination.

1. WHO, 2009: WHO Guidelines on Hand Hygiene in Healthcare: a summary 2. Folkehelseinstituttet, 2004: Nasjonal Veileder for handhygiene3. Hughes KA, Cornwall J, Theis J-C, Brooks HJL. Bacterial contamination of unused, disposable non-sterile gloves on a hospital orthopaedic ward, 2013 4. Diaz et al, 2008: Contamination of examination gloves in patient rooms and implications for transmission of antimicrobial-resistant microorganisms 5. Swann-Morton Study, 2009: Six weeks trial with SafeDon 6. Swann-Morton Study, 2010: Six weeks trial with SafeDon

• SafeDon Hygiene System dispense one glove at the time and eliminates touching of the opening and other "unused" gloves inside

• SafeDon Hygiene System eliminates unwanted gloves from falling out of the box and therefore reduces waste



• Resists permeation by a wider range of chemicals than natural rubber latex of the same thickness

• 100% nitrile material eliminates Type I allergic reactions associated with NR-Latex protein

• Superior strength and puncture resistance

• SafeDon Hygiene System eliminates users from touching the most critical surfaces of the glove: - thumb, fingers and palm area

PRODUCT INFORMATION

Type: Powder free and non-sterile

Primary Material: Acrylonitrile - butadiene

Powder: No powder lubricant added

Color: White

Design and Feature: Ambidextrous, textured surface at fingers and beaded cuff

Packing: 200 gloves per dispenser (180 XL)

Dimensions:

Glove Size	Reorder#	Palm Width (mm)	Length (mm)
Extra-small	SDN-100-01	≤ 80	240 min
Small	SDN-100-02	80 ± 10	240 min
Medium	SDN-100-03	95 ± 10	240 min
Large	SDN-100-04	110 ± 10	240 min
Extra-Large	SDN-100-05	≥ 110	240 min

Thickness:

Location of Thickness Measurements		Single Wall (mm)
Finger	(at 15 mm from the extreme tip)	min. 0.07
Palm	(at center of palm)	min. 0.06
Cuff	(at 25 mm from cuff end)	min. 0.05

Physical Properties:

Parameters	Before Aging	After Aging
Tensile Strength (MPa)	≥ 14	≥ 14
Force At Break (N)	≥ 6	≥ 6

Pre-shipment Quality Inspection:


Criteria	Insp. Level	AQL
Dimension	N=13	Median
Physical Properties	N=13	Median
1000ml Water Leak Test	G-1	1.5
Visual Inspections for Major defects	G-1	2.5
Visual Inspections for Minor defects	G-1	4.0
Powder Free Residue	N=5	N/A

Product Conformance:

- Medial Device: in compliance with European Medical Directive 93/42/EEC (CE Class I)
- EN 455
- ASTM D6319

Quality Assurance:

- US FDA Quality System Regulation (QSR)
- ISO9001 Quality System
- ISO13485 Quality Management System



SafeDon™
Hygiene System

www.safedon.com
post@safedon.com

REMESCO Handelsges. m.b.H
Grinzinger Allee 5/25, A-1190 Vienna,
Austria
Office: +43-1-328-50-88
Fax: +43-1-328-29-49
Email: office@remesco.com

Distributed by: